



Government of West Bengal
Office of the Chief Medical Officer of Health
Department of Health & Family Welfare
South 24 Parganas

Administrative Building, M R Bangur Hospital Complex, 241, Deshapran Sashmal Road, Tollygunge, Kolkata 33.
Mail Id- cmohs24pgs@gmail.com, Website- www.spghealthgov.in, Phone no / Fax - 033-2422 1037, 033-2422-0124

Memo No: CMOH/ (SPG)/Tender/ 12319


Date: 17.12.2020

Tender Notice for Procurement
Non-Sterile Surgical Rubber Gloves

District Health & Family Welfare Samity (DH & FWS) South 24 Parganas invites quotation from reputed company/firm/agency/for supply of the following items of "**Non-Sterile Surgical Rubber Gloves**" list attached (Annexure - I) as per specification available DH & FWS cum CMOH , South 24 Parganas Office on any working day during office hours.

Terms & Conditions & Require Documents:

1. Tender will be submitted from office of the CMOH, South 24 Parganas under submission of application in a letter head of the concern provider during office hours from 17.12.2020 to 23.12.2020 at 12 PM. One bidder is eligible to submit one bid for indented item(s) only.
2. Work to be must confirm to the specification as given above. If any deviation is found, the department reserves the right to reject the materials at its sole discretion.
3. Self attested Copy of (Bidder) PAN Card, Trade Licence (Medical & Equipment) and GST registration certificate are to be attached with quotation.
4. Payment shall be made after completion the work within specified time period.
5. The rate should be quoted in total value excluding GST.
6. In case of tie or rates, draw of lots will be done.
7. Sample must be submitted during the submission of Bid documents (mandatory)
8. After selection of Agency Test report will must be submitted during supply.
9. The last date for submission of Tender paper: 23th 12. 2020 upto 12 P.M at Drop Box CMOH Office, South 24 Parganas, CMOH office, South 24 parganas.
10. Tender will be opened on 23th 12. 2020 at 2 P.M at CMOH Office Chamber.
11. The Chief Medical Officer of Health, South 24 Parganas shall reserve the right cancellation, adding reducing or deferring the Tender in total or in partial without assigning any reason thereof.


CMOH, Secretary
District Health & Family Welfare Samity,
South 24 Parganas

Annexure – I)

Specification of Non-Sterile Surgical Rubber Gloves

1. Make RRTI-Anchor/Jacob Scientific- Surgicare/ Paramount – Surgimed / Medi-Grip / Midas/ Bodyguard equivalent conforms to IS-4148 in total.
2. Materials shall be made by the dipping process using Natural rubber Latex finish of outer surface shall be smooth. The gloves shall be transparent or translucent.
3. The gloves shall not contain any ingredients known to be harmful either to the wearer or to persons they may come in contact with.
4. The gloves shall beaded cuff with a bead of not more than 3.00 mm and shall have a rolled wrist.
5. The Gloves shall be free from visible imperfection.
6. Thickness as per IS 4148: 1989 i.e. 0.24+/-0.06mm (measured by dead weight dial type gauge with foot pressure of 20kPa on the rubber)
7. Lightly powdered with absorbable dusting powder.
8. Can withstand repeated autoclaving cycles or temperature up to 70C
9. Marking as per IS 4148 (optional)
10. Tensile strength : Min 18Mpa & Elongation at break Min-700%
11. Bidder shall indicate make of them or compliance of IS 4148
12. Pre- Dispatch Inspection is not involved however items received will be checked randomly for dimensions, thickness, materials, visual checks etc. Tests may be conducted by purchaser on random sample selected from total lot (bulk supply) at govt approved lab. Total lot will be rejected based on test results.

Rate of Item

Item Name

Rate per
(Pack of 1 pair x 25)

➤ **Latex Gloves Non-Sterile
Surgical Rubber
Size No- 6, 6 ½, 7, 7½, 8**

Full Signature of the bidder/
Authorised person & Seal